

United States Patent and Trademark Office



ATTORNEY DOCKET NO. APPLICATION NO. FILING DATE FIRST NAMED INVENTOR CONFIRMATION NO. 7175 BET I-115 CIP 10/670,104 09/24/2003 John B. Fenning EXAMINER 1218 7590 05/19/2005 REIMERS, ANNETTE R CASELLA & HESPOS 274 MADISON AVENUE ART UNIT PAPER NUMBER NEW YORK, NY 10016 3732

DATE MAILED: 05/19/2005

Please find below and/or attached an Office communication concerning this application or proceeding.

	<u></u>	
•	Application No.	Applicant(s)
Office Action Summary	10/670,104	FENNING ET AL.
	Examiner	Art Unit
	Annette R. Reimers	3732
The MAILING DATE of this communication appears on the cover sheet with the correspondence address Period for Reply		
A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION. - Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication. - If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely. - If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication. - Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).		
Status		
1) Responsive to communication(s) filed on		
2a) This action is FINAL . 2b) ☑ Thi	is action is non-final.	
3) Since this application is in condition for allowance except for formal matters, prosecution as to the merits is		
closed in accordance with the practice under Ex parte Quayle, 1935 C.D. 11, 453 O.G. 213.		
Disposition of Claims		
4) Claim(s) <u>1-9</u> is/are pending in the application.		
4a) Of the above claim(s) is/are withdrawn from consideration.		
5) Claim(s) is/are allowed.		
6) Claim(s) <u>1-4 and 6-9</u> is/are rejected.		
7) Claim(s) 5 is/are objected to.		
8) Claim(s) are subject to restriction and/or election requirement.		
Application Papers		
9) The specification is objected to by the Examin	er.	
10)⊠ The drawing(s) filed on <u>24 September 2003</u> is/are: a) accepted or b)⊠ objected to by the Examiner.		
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).		
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).		
11) The oath or declaration is objected to by the E	examiner. Note the attached Office	Action or form PTO-152.
Priority under 35 U.S.C. § 119		
12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a) All b) Some * c) None of:		
1. Certified copies of the priority documents have been received.		
2. Certified copies of the priority documents have been received in Application No		
3. Copies of the certified copies of the priority documents have been received in this National Stage		
application from the International Bureau (PCT Rule 17.2(a)).		
* See the attached detailed Office action for a list of the certified copies not received.		
Attachment(s)		
 Notice of References Cited (PTO-892) Notice of Draftsperson's Patent Drawing Review (PTO-948) Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08 Paper No(s)/Mail Date 09/24/03, 01/12/04. 	4) Interview Summary Paper No(s)/Mail Do 5) Notice of Informal F 6) Other:	

DETAILED ACTION

Drawings

The drawings are objected to as failing to comply with 37 CFR 1.84(p)(5), because they do not include the following reference sign(s) mentioned in the description: reference number 110 on page 15, line 13, of the specification. Corrected drawing sheets in compliance with 37 CFR 1.121(d) are required in reply to the Office action to avoid abandonment of the application. Any amended replacement drawing sheet should include all of the figures appearing on the immediate prior version of the sheet, even if only one figure is being amended. Each drawing sheet submitted after the filing date of an application must be labeled in the top margin as either "Replacement Sheet" or "New Sheet" pursuant to 37 CFR 1.121(d). If the changes are not accepted by the examiner, the applicant will be notified and informed of any required corrective action in the next Office action. The objection to the drawings will not be held in abeyance.

Claim Rejections - 35 USC § 112

The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.

Claims 1-9 are rejected under 35 U.S.C. 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter, which the applicant regards as the invention.

In claim 1, lines 10-12, "the first notch" and "said second notch" lack proper antecedent basis.

In claim 7, line 11, "the posterior control arm notch" lacks proper antecedent basis.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1-3 and 6-8 are rejected under 35 U.S.C. 102(b) as being anticipated by Pappas (U.S. Patent Publication Number 2001/0034555).

Pappas discloses a knee joint prosthesis comprising a tibial component, 500, with a superior bearing surface, a bearing, 300, having an inferior surface in sliding bearing engagement with the superior surface of the tibial component (see Figures 1), a groove extending substantially from anterior to posterior in the inferior surface of the bearing (see Figure 4), and a recess formed in the inferior surface of the bearing at an end of the groove adjacent the anterior extreme of the bearing (see Figure 5).

The device further comprises a control arm, 410, having a pivotal support, 430, pivotally mounted to the tibial component and slideably engaged in the groove of the bearing, an anterior control arm stop pin, 417, mounted to the anterior end of the control arm and engaged in the anterior notch of the bearing, and a posterior control arm stop pin, 411, extending from the control arm and engaged in the posterior notch of the bearing (see Figures 4 and 7). The prosthesis further comprises a rotation-limiting stop pin, 420, fixed relative to the tibial component (see Figure 7).

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.

Claims 4 and 9 are rejected under 35 U.S.C. 103(a) as being unpatentable over Pappas (U.S. Patent Publication Number 2001/0034555).

Pappas discloses the claimed invention except for the rotation-limiting stop pin configured relative to the second component for limiting rotational movement to approximately 30 degrees and the rotation-limiting stop and the bearing configured for permitting approximately 15 degrees of rotation of the bearing in either direction from a central position. It would have been obvious to one having ordinary skill in the art at the time the invention was made to construct the device of Pappas with the rotation-limiting stop pin configured relative to the second component for limiting rotational movement to approximately 30 degrees and the rotation-limiting stop and the bearing configured for permitting approximately 15 degrees of rotation of the bearing in either direction from a central position, since it has been held that where the general conditions of a claim are disclosed in the prior art, discovering the optimum or workable ranges involves only routine skill in the art. In re Aller, 105 USPQ 233.

Allowable Subject Matter

Claim 5 is objected to as being dependent upon a rejected base claim, but would be allowable if rewritten in independent form including all of the limitations of the base claim and any intervening claims.

Conclusion

The prior art made of record and not relied upon is considered pertinent to applicant's disclosure. See PTO 892 for art cited of interest.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Annette R. Reimers whose telephone number is (571) 272-7135. The examiner can normally be reached on Monday-Friday.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Kevin Shaver can be reached on (571) 272-4720. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

AR

W

SUPERVISORY PATENT EXAMINER TECHNOLOGY CENTER 3700